

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kodikaral	Details About Robust Analysis								
Lab No	10555	Detail About Monthly Summary								
Month	June	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data		All Analyser Result	Print							
		Date of Result Entered : 20/06/2023								
		Date of Report Published : 04/07/2023								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	275 mg/dL	-0.78	0.69
2	UREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	13	2.55	7.09	0.18	2.6 mg/dL	0.28	0.10
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	118 mg/dL	0.84	0.24
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out