

			CAL	<u>IBRATIO</u>	N CER	<u> I IFICAT</u>	<u>E</u>		
CERTIFICATE NO: SBS/CL/23/13604				MECHANIC		P	age No:1 of 1		
Issue Date				07-10-2023					
SRF No & Date				SRF/	SRF/23/00776-0005 & 06-10-2023				
Receipt Date				06-10	06-10-2023				
Calibration Date				06-10	06-10-2023				
Calibra	ation Due			05-10	05-10-2024				
Custo	mer Name	& Add	fress						
GOVE	RNMENT F	PRIMA	RY HEALTH CENTRE,						
KODIK	KARAI-614	4807,N	AGAPATTINAM DISTR	ICT.					
				etails of Device	Under Calibr	ration (DUC)			
Descri	ption	:	CENTRIFUGE	Make	& Model	:	REMI & C-852		
Range : 5250 RPM		Sr. No		:	ZFAN-00504				
Resolution : 10 RPM		10 RPM	Identification No		:	NA			
DUC Condition : Satisfactory		Locat	Location		LABORATORY				
			Environmenta	I Conditions & S	Standard Ope	rating Proced	ure Details		
Environmental Details Temperatur				ure: 24.3 °C	re: 24.3 °C Relative H		54 % Rh		
Calibration Procedure No SBS/CP/ML				/IL/04	_/04 Calibratio		n done at ONSITE		
			1	Reference	Standards De	etails			
S.No	o Description			Make/ SI No:	Make/ SI No:		Certificate No		
-	Digital Tachometer			Lutron & DT-2238		JRPM-CCTR-A&S-2022-0036		Validity 03-10-2023	

DEVICE LINDED

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.9	0.1	4.2
2	2000	1998	2	4.2
3	3000	2999	1	4.2

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

P.PRASANNA

MEDI Chennai 600 032

(C.SIVABALAN)

Authorised Signatory

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

