

				CA	LIBRA	4110	N CEF	<u> (</u>	- <i>IC</i>	AIE	
CERTIFICATE NO: SBS/CL/23/13603				MEDICAL DEVICES					Page No:1 of 1		
Issue Date				07-10	07-10-2023						
SRF No & Date				SRF/	SRF/23/00776-0004 & 06-10-2023						
Receipt Date				06-10	06-10-2023						
Calibration Date				06-10	06-10-2023						
	ation Due					05-10	05-10-2024				
Custo	mer Nam	e & Ad	dress								
GOV	ERNMENT	PRIMA	ARY HEA	ALTH CENT	RE,						
KODI	KKARAI-6	14807,1	NAGAPA	TTINAM DIS	TRICT.						
					Details o	f Device	Under Calib	ration	(DUC	()	
Description : (MICROSCOPE)			TY	Make & Model : LAWERNCE&MAYO & XSZ-N1071			& XSZ-N1071				
Range		1	MULTI			Sr. No	0		:	NA	
Resol	ution	1	MULTI			Identi	fication No		:	NA	
DUC (Condition	1	SATISI	FACTORY		Locat			:	LABORATORY	
						onditions	& Calibrati	on Pr	ocedu	re Details	
Environmental Details Temperature:25.6*			:25.6°C	5.6°C Relative Hu		midity 52% RH		52% RH			
Calibration Procedure No SBS/CP/MD/			29 Calibration		Calibration of	done at ON:		ONSITE			
					Re	eference	Standards [Details			
S.No	No Description			Make/ SI No:		Certificate No			Validity		
	Electrical Safety Analyser			Rigel Medical & 44L-1059		M-230809-16-4					

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)	
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MΩ	97	13.92	
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)	
	<5000µAfor B,BF,CF	125	12.3	
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	229	8.6	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By Authorised Signatory





_	- 101 2
	Techinical Manager
_	C SHANMUGARA I

Chief	Executive

SUNSHINE BIOMEDICAL SOLUTIONS