



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name Govt. PHC, Kuttapatti Lab No 12593  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based Date of Report Published 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.84	A
2	CREATININE III	Group I	1927	0.93	A
3	CHOLESTROL III	Group I	1923	0.64	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ±3.00	D	Unacceptable

LAB ADDRESS  
Govt. PHC, Kuttapatti  
Kuttapatti  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College,  
Vellore

Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

OMSDI Range	Grade
0.00 - 1.00	A
1.01 - 2.00	B
2.01 - 2.99	C
Beyond ±3.00	D

LAB ADDRESS  
Govt. PHC, Kuttapatti  
Kuttapatti  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact number: 0416-2283102