

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC ,Vedaranyam	Details About Robust Analysis								
Lab No	12424	Detail About Monthly Summary								
Month	April	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
Click on the analyte to view Graphical Data										
		All Analyser Result	Print							
Date of Result Entered : 19/04/2023										
Date of Report Published : 06/05/2023										
Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119.6 mg/dL	0.02	0.34
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	13	2.20	10.23	0.22	2.37 mg/dL	0.76	0.12
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	102.4 mg/dL	-1.59	0.28
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out