

## CERTIFICATE OF CALIBRATION

<b>Certificate No</b>	CAL-JK-Z2-2023-10-1019		
<b>Date of Receipt</b>	28-09-2023	<b>Issue Date</b>	29-09-2023

<b>Issued By</b>	Trimed Solutions (I) Private Limited		
<b>Address</b>	2nd Floor, Jyothi Elegance, Kavuri Hills, Phase-III, Madhapur, Hyderabad-500081		

<b>Health Facility</b>	Srinagar (JLNM) DH		
<b>State</b>	Jammu & Kashmir	<b>District</b>	Srinagar

### Medical Equipment Details

<b>ME Number</b>	JKZ2SRIDH0200998	<b>Manufacturer</b>	Whirpool
<b>ME Name</b>	Refrigerators, Laboratory (Lab refrigerator)	<b>Model</b>	DC 205F (71622)
<b>Department</b>	ICTC LAB	<b>Serial No</b>	INC202209542
<b>Range</b>	2°C to 8°C	<b>Resolution</b>	0.1°C
<b>Calibrated at</b>	On site	<b>Condition of Equipment</b>	Visually good
<b>Date of Calibration</b>	28-09-2023	<b>Calibration Due</b>	27-09-2024

### Environmental Conditions

<b>Temperature in °C</b>	22°C	<b>Relative Humidity in % Rh</b>	40%
--------------------------	------	----------------------------------	-----

### Details of Master Equipment used for Calibration

No	Description	Equipment ID	Certificate Number	Date of Calibration	Calibration Due
1	Digital Thermometer	TRI-TE-0296	23.01.05.017	05-01-2023	04-01-2024

### Calibration Standards

<b>Applicable Specification</b>	By Using Digital Thermometer By Comparison Method	<b>Calibration Procedure</b>	TSPL-CAL-06
---------------------------------	---	------------------------------	-------------



**Performance Result**

No	Parameter	Unit	Set value	Measured Value	Acceptable range	Remarks
1	Temperature	°C	2 To 8	2	1.5 To 8.5	Pass

**Conditions of Issue of Report**

1. The standards used are traceable to National and International Standards through unbroken chain of Accredited Laboratories.
2. Results reported are valid at the time of & under the stated condition of the measurements.
3. Any error in the certificate should be brought to our knowledge within 30 days from the date of the certificate.
4. This calibration report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.
5. Decision Rule - Simple Acceptance where "PASS" indicates measured value is within/Equal to the Tolerance mentioned by Manufacturer specification and "FAIL" indicates measured value is out of Tolerance without considering the measurement Uncertainty.

<b>Calibration Status</b>	<b>Pass</b>	Measured Values are within the Specified Limit
---------------------------	-------------	--

<b>Calibrated By: Aijaz Gulam Kumar</b> <b>(Calibration BME)</b> 	<b>Authorised Signatory: Sonu Rajput (Technical Manager)</b> 
--	--