



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	
Calibration Date: 28.09.2023	
Calibration valid: 27.03.2024	
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CENTRE, ERNAPURAM, NAMAKKAL.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	MICROLUX
ii Model	-----
iii Value	100µl-1000µl
iv Serial No	UPHC/EM/PIP/002

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023





Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)

