

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Subbalapuram.M		<a href="#">Details About Robust Analysis</a>							
Lab No	7216		<a href="#">Detail About Monthly Summary</a>							
Month	August		<a href="#">Detail about SDI</a>							
Year	2023									
Constituent Group	Chemistry III									
<a href="#">Click on the analyte to view Graphical Data</a>			<a href="#">All Analyser Result</a>		<a href="#">Print</a>					
			Date of Result Entered : 19/08/2023							
			Date of Report Published : 01/09/2023							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	173 mg/dL	-0.29	0.42
2	<a href="#">UREA III</a>	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	33 mg/dL	1.10	0.65
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	2.1 mg/dL	2.86	0.01
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	5.43	10.18	0.55	4.5 mg/dL	-1.68	0.17
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	93 mg/dL	-4.48	0.22
6	<a href="#">TRIGLYCERIDE III</a>	GPO-PAP / Enzymatic Colorimetric / End Point III	Robonik ( Semi automatic / Fully Automatic )	11	137.31	8.60	11.80	167 mg/dL	2.52	7.12
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>