

AUGUST - 2023.

1. Name of the Test	Name of the Test Method	Name of the kit & Batch no Expiry	Values mg/dl.	Report updated	SPI	Interpretation
1. Glucose	GOD-POD	Glucose Lot - 047220610 Ex 4/24	173 mg/dl	19.8.23	-0.29	Excellent
2. Urea	GLPH	Urea. Lot - 117220914 Ex 2/24	28 mg/dl	"	-0.07	Excellent
3. Creatinine	Jaffes. Colorim.	Creatinine Lot CMR-010123 Ex - 6/24.	1.15 mg/dl.	"	-0.08	Excellent.

(Signature)
Sany