

JULY - 2023

S. No	Name of the Test	Test Method	Name of the kit Batch No & Expiry	Values mg/dl.	Report updated	SDI	Interpretation
1.	Glucose	GOD. POD	Glucose Lot - 047220610 Ex 2/24	113 mg/dl	19.7.23	0.60	Excellent
2.	Urea	GLPH.	Urea Lot - 117220914 Ex 2/24	71 mg/dl	"	-0.07	Excellent
3	Creatinine	Jaffe's kinetic	Creatinine Lot-CMR - 010123 Ex - 6/24	7.8 mg/dl	"	-0.07	Excellent
4	Cholesterol	CHOD-POD	Cholesterol chol - 040622 1/23	110 mg/dl.	"	-0.64	Excellent

Singh