

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Munichalai	<a href="#">Details About Robust Analysis</a>									
Lab No	10786	<a href="#">Detail About Monthly Summary</a>									
Month	July	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 19/07/2023									
		Date of Report Published : 07/08/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	115 mg/dL	0.93	0.28	
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.9 mg/dL	3.06	0.02	
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>