

## G MULTI SERVIC

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452. gmultiservicesgobi@gmail.com. Cell: 96597 20654. **Bio Medical Instruments Services** 

Ey. A. Ibathul Raziq DECE., BE (ICE) **Eio Medical Engineer** 

## Calibration Protocol & Report for Medical Equipment

Equipment: PIPETTE

: FINNPIPETTE

MAKE MODEL

:100-1000µL

PROTECT CLASS / TYPE : Risk Class II

S. No.

: RW11700

Location

: GPHC, SENTHANADU

S.No	Me	easurement Parameters	Status	Remarks
1	Electrical Test	(a)Safety Standards (range: 210-240 V) (b)Equipotential ground & Earthing (c)Risk of shock/fire Hazards,if any (d)Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status		-
2	Physical	(a)Cabinet & Visible damages, if any	No	-
	Condition Test	(b) Risk of mechanical damages, if any	No	Checked OK
3	Performance Test	(e) Mean 99.2µl	ОК	
,		(f) A -0.5 μI (g) Acc % -0.71 (h) CV% -0.36	OK OK Ok	
1	Calibration Test	min volume 100 μl max Volume 1000 μl (ELECTRONIC BALANCEING TEST) (e) A+or - 1.5 , 5 μl (f) S 1.5 μl max Vol. 1000 μl (g) Acc+ or - 1.50, 0.50 (h) ÇV% 0.50,0.10	Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:	User Department:	
Research team head  Reviewed On: 20.09.23  Next Review: 19.09.24		* GOBI *		