


CENTRIFUGE Calibration / Verification Record Sheet				
Location: PPTCT-MCH-ADONI-KURNOOL			Range : Up to 3500RPM	
Equipment ID: APSACS/PPTCT/MCH/ADN/CEF/01			SL.NO :REMI - LOLC-10259	
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer				
Date of Calibration :	10.10.2023		Due date of Calibration:	09.10.2024
Tachometer ID : KMC/MB/HIVTL/SRL/TM/01			Calibration date : 19.04.2023	Due date of Calibration : 18.04.2024
Verification Done by: B.BHASKAR, Technical Officer, SRL, KMC,KURNOOL				
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)	
1.	499	500	1.0	
2.	993.9	1000	7.1	
3.	1997.8	2000	2.2	
4	2998.2	3000	2.8	
5.	3497.1	3500	2.9	
Accept/ Reject: Accepted since average difference is within +/- 100RPM				
Final Interpretation: Centrifuge is in good working condition				
Action Taken: - NA				
Remark: - NA				
				 TECHNICAL OFFICER