

BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

| Site name: | DISTRICT HOSPITAL SATNA |
|-----------------------|--|
| | Sardar Vallabh Bhai Patel District Govt. Hospital, Jagat dev Talab Road Pushpraj Colony Satna (M.P) |
| Province- Postal code | 462039 |
| Contact Name | Mr. Anshul Soni |

Calibration Certificate

| Date | 13.08.2023 | Instrument | D-10Hb Analyzer | Instrument | DM0L566540 |
|------|-----------------|------------|--------------------|---------------|------------|
| | Thytotoman 2007 | Model | B 10118 / many 201 | Serial Number | |

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

| Measured with (Device manufacturer and model | Device Serial Number | Device Calibration date (dd,mm,yyyy) |
|--|----------------------|--------------------------------------|
| Measuring cylinder hex base 10ML | 2310100 | 20.12.23 |

| Specified Range@1.5ml/Minute | Measured Value |
|------------------------------|----------------|
| 1.425 - 1.575ml/minute | 1.49 |

Column Holder temperature Reference: Service Manual: Section 7.9

| Measured with (Device manufacturer and model | Device Serial Number | Device Calibration date (dd,mm,yyyy) |
|--|----------------------|--------------------------------------|
| FLUKE179 | 26030078 | 01.12.24 |
| Specified Range@35°C | Measured Value | |
| 33.4°C – 36.4°C | 34.9 | |

Detector log value

Reference: Service Manual: Section 7.2 or Field Bulletin D-10 - 081

| Measured with | Specified Range | Measured Value |
|-----------------------------|-----------------|----------------|
| Instrument Service Software | 0.018 - 0.022AU | 0.021AU |

The next calibration is due on 12.08.2024 (dd,mm,yyyy)

For Bio-Rad Laboratories India Pwt Limited

Authorized Signatory: Sarlu

Page1of1

Calibration Certificate