

9/27/23, 3:05 PM



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name: Govt. Block PHC, Perungattur
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7533
Date of Result Entered: 19/08/2023
Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	188 mg/dL	1.38	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	128 mg/dL	3.06	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Perungattur
THE BLOCK MEDICAL OFFICER, GOVT PHC , PERUNGATTUR , CHEYYAR TALUK . THIRUVANNAMALAI DISTRICT.
Perungattur , Cheyyar Tk.
TAMILNADU604402

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Cor.tact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

O. R. Seen
MEDICAL OFFICER
GOVT PRIMARY HEALTH CENTRE
PERUNGATTUR 604 402
VEMBAKKAM TK T V M JAI DI