

VIEW LAB MONTHLY SUMMARY

| | | |
|-------------------|----------------------|-------------------------------|
| Lab Name | Govt.UGPHC, Arimalam | Details About Robust Analysis |
| Lab No | 7147 | Detail About Monthly Summary |
| Month | February | Detail about SDI |
| Year | 2023 | |
| Constituent Group | Chemistry III | |

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/02/2023


Date of Report Published : 08/03/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1868 | 155.11 | 6.56 | 10.18 | 150 mg/dL | -0.50 | 0.47 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1872 | 3.40 | 9.00 | 0.31 | 3.6 mg/dL | 0.65 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1788 | 112.79 | 5.23 | 5.90 | 110 mg/dL | -0.47 | 0.28 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |


 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE,
 ARIMALAM - 622 201,
 PUDUKKOTTAI DISTRICT