

Microlab 300 Mahabuli
Name & Address of Customer : Green Park Diagnostics
G-43, Green Park, Main Market
Delhi
STATE : Delhi PIN : 110016
Phone : 9582859222 E-Mail : Greenpark43@yahoo.co.in
Name of Instrument : **Microlab ³⁰⁰ LX**(w/o inbuilt printer)
Microlab ³⁰⁰ EX(with inbuilt printer)
Type : Semi Automatic Biochemistry Analyser
Serial No : 0-0811
Installation Date : 15/Oct/2015
Calibration Date : 05/Oct/2023
Next Calibration Due on : 04/Oct/2024

This is to certify that above said instrument has been validated by hardware calibration for Filters, Aspiration (Pump), Temperature & Lamp according to the procedures provided by Vital Scientific, Netherlands (Elitech Group).

This calibration is carried out by using Standard Operating Procedures (S.O.P.) provided by Vital Scientific (Elitech Group Company) and traceable to Vital Scientific standards.

This instrument conform to IVD & EU directives of use.

Calibration carried out on site by:

Ganesh Singh

Signature

Name of Service Engineer :
Instrument Service Division
Q-Line Biotech Pvt. Ltd.

Ganesh Singh

Q-LINE BIOTECH PRIVATE LIMITED (Formerly known as "POCT Services Pvt. Ltd.")

Corp. Office: 20A (Basement), Main Shivaji Marg, Najafgarh Road, Near Moti Nagar Police Station, New Delhi-110015, INDIA. Tel.: 011-45577407
Regd. Office: 298-281, Transport Nagar, Kanpur Road, Adjacent Transport Nagar Metro Station, Lucknow-226023 (U.P.), INDIA. Tel.: 0522-2433023

E-mail: info@qlinebiotech.com | Website: www.qlinebiotech.com

12-01-2022

Name & Address of Customer :

Green Park Diagnostics
G-43, Green Park, Main Market
Delhi

STATE: Delhi PIN: 110016

Phone:

9582859223

E-Mail :

Greenpark43@yahoo.co.uk

Name of Instrument : Microlab 300 LX(w/o inbuilt printer) Microlab 300 EX(with inbuilt printer)

Type : Semi Automatic Biochemistry Analyser

Serial No :

17-46389

Installation Date :

07/07/2018 / 2018

Calibration Date :

09/04/2023

Next Calibration Due on

04/04/2024

This is to certify that above said instrument has been validated by hardware calibration for Filters, Aspiration (Pump), Temperature & Lamp according to the procedures provided by Vital Scientific, Netherlands (Elitech Group).

This calibration is carried out by using Standard Operating Procedures (S.O.P.) provided by Vital Scientific (Elitech Group Company) and traceable to Vital Scientific standards.

This instrument conform to IVD & EU directives of use.

Calibration carried out on site by:

Gantabulchi

Signature

Name of Service Engineer :

Instrument Service Division

Q-Line Biotech Pvt. Ltd.

Q-LINE BIOTECH PRIVATE LIMITED (Formerly known as "POCT Services Pvt. Ltd.")Corp. Office: 20A (Basement), Main Shivaji Marg, Najafgarh Road, Near Moti Nagar Police Station, New Delhi-110015, INDIA. Tel.: 011-45577407
Regd. Office: 298-281, Transport Nagar, Kanpur Road, Adjacent Transport Nagar Metro Station, Lucknow-226023 (U.P.), INDIA. Tel.: 0522-2433023

E-mail: info@qlinebiotech.com | Website: www.qlinebiotech.com

12-01-2022