

Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area Delhi - 110092



Phone No.: 09810943334 , 011-41764006, Email: multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web.: www.multitechtechnofab.com

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20053 ULR: CC360323000002330F Calibration Issued on: 27-09-2023

Job / Identification No. : MTE/CAL/JOB/22356 Date of Calibration : 18-09-2023

Item Receipt Date: 18-09-2023 Recommended Date of Next Calibration: 18-09-2024

Calibrated For M/S Green Parq Diagnostics Reference : Order thru Telephonic

G43 Green Park, Main Market, New Delhi -16 P. O. No. : N/A

Discipline Mechanical Group Acceleration & Speed

Certificate for : Laboratory Centrifuge Item Location : NA

Make / Model	Serial / I.D. No.	Inst. Range	Least Count	Unit	Item Condition
Remi / R-8C	NA / NA	0 to 5250	0.0	RPM	Working
Environmental Condition		24.70 ± 0.5 °C		52 ± 3% RH	

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make / Model	Certificate No.	Calibration Agency	Valid up to
Digital Tachometer(Non contact)	Lutron / DT-2234C BCL/2023/0166F		Bhagwati Calibration Laboratory	17-01-2024

Lab Procedure: MTE-2/RPM/2.50

RESULTS: On-Site Calibration

All Readings are in RPM, Specified Otherwise.

Calibration Procedure

U.U.C. READING (SET AT)	AVERAGE	UNCERTAINTY^ in (%)
700	700.9	± 2.31
1400	1403.0	± 2.31
2100	2103.0	± 2.31
2800	2803.0	± 2.31
3500	3503.0	± 2.31

^The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

For and on behalf of Multitech Technofab Enterprises

STD Procedure: NA

Calibrated By DEEPAK KUMAR Technical Executive

Laboratory Head



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area Delhi - 110092

Phone No.: 09810943334, 011-41764006, Email: multitech.technofab@gmail.com, calibration@multitechtechnofab.com, Web.: www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20053

ULR: CC360323000002330F

Calibration Issued on: 27-09-2023

• The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
 The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor
- 2. The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- 3. Total liability of our organization is limited to the invoiced amount.
- 4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- 5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

For and on behalf of Multitech Technofab Enterprises

Calibrated By DEEPAK KUMAR Technical Executive

Laboratory Head