



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com



CC3603

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20049

ULR :
CC360323000002334F

Calibration Issued on: 27-09-2023

Job / Identification No. : MTE/CAL/JOB/22355

Date of Calibration : 18-09-2023

Item Receipt Date: 18-09-2023

Recommended Date of Next Calibration: 18-09-2024

Calibrated For

M/S Green Parq Diagnostics
G43 Green Park, Main Market, New Delhi -16

Reference : Order thru Telephonic
P. O. No. : N/A

Discipline

Thermal

Group

Temperature

Certificate for : Temperature for Incubator

Item Location : NA

Make / Model

Serial / I.D. No.

Inst. Range

Least Count

Unit

Item Condition

NA / NA

NA / NA

Mentioned
hereunder

As Per Range

°C

Working

Environmental Condition

25.35 ± 0.5 °C

52 ± 3% RH

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard

Make / Model

Certificate No.

Calibration Agency

Valid up
to

4 Wire RTD (PT-100) Sensor
with Indicator

Tempens and Lutron / PT-100
and TM-917

ULR-
CC288822000002063F

Sigma Test & Research
Centre

13-12-
2023

Calibration Procedure

Lab Procedure : MTE-2/T/2.5

STD Procedure : DKD-R-5-4 & DKD-R-5-7

RESULTS: On-Site Calibration

All Readings are in °C, Specified Otherwise.

LOCATION

CH-1

Temperature

Range 0 to 70 °C

30

30.09 °C

UNCERTAINTY[^] °C : 1 °C

35

35.08 °C

UNCERTAINTY[^] °C : 1 °C

37

37.08 °C

For and on behalf of
Multitech Technofab Enterprises

Calibrated By DEEPAK KUMAR
Technical Executive

Laboratory Head



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20049

ULR :
CC360323000002334F

Calibration Issued on: 27-09-2023

LOCATION

CH-1

UNCERTAINTY[^] °C : 1 °C

40

40.08 °C

UNCERTAINTY[^] °C : 1 °C

45

45.08 °C

UNCERTAINTY[^] °C : 1 °C

^The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.
- CH Stands for Channel.

Note

1. This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
2. The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

For and on behalf of
Multitech Technofab Enterprises

Calibrated By DEEPAK KUMAR
Technical Executive

Laboratory Head