

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 04.10.2023
		Calibration valid : 03.04.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , SHEVAPET, SALEM CORPORATION, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINN PIPETTE F3
iii	Value	100µl-1000µl
iv	Serial No	RW11520



Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.01	500.02	1000.05
2	100.02	500.01	1000.04
3	100.00	500.02	1000.02
4	100.02	500.02	1000.04
5	100.01	500.05	1000.05
6	100.05	500.02	1000.04

For Heart Beat Medical Systems,


(Authorized signatory)


MD Reg No: TN/CBE/MD42/00016


GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 04.10.2023
		Calibration valid : 03.04.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , SHEVAPET, SALEM CORPORATION, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINNPIPETTE – F3
iii	Value	10 µl -100 µl
iv	Serial No	RW09114

Calibration standards used :


Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.01	100.03
2	10.02	50.04	100.02
3	10.01	50.01	100.02
4	10.03	50.02	100.05
5	10.00	50.05	100.04
6	10.02	50.03	100.02

For Heart Beat Medical Systems,

(Signature)
(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navaindia Road, Peelamedu,
CBE – 04, hbmedicalsystem@gmail.com