

CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 04.10.2023
		Calibration valid : 03.04.2024
1.1. CUSTOMER DETAILS:		
Name and Address of the organization		M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , SHEVAPET, SALEM CORPORATION, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	REMI
ii	Model	R-SC
iii	Serial No	UPHC/SP/CENT/001

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	2005	5

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016


GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2



Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by 


MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
88833 15662

No.42/1, Navaindia Road, Peelamedu,
CBE - 04, hbmedicalsystem@gmail.com