



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com

MULTI TECH
Technofab Enterprises

CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20052

ULR :
CC360323000002331F

Calibration Issued on: 27-09-2023

Job / Identification No. : MTE/CAL/JOB/22350

Date of Calibration : 20-09-2023

Item Receipt Date: 18-09-2023

Recommended Date of Next Calibration: 20-09-2024

Calibrated For

M/S Green Parq Diagnostics
G43 Green Park, Main Market, New Delhi -16

Reference : Order thru Telephonic
P. O. No. : N/A

Discipline

Electro-Technical

Group

Time & Frequency

Certificate for : Timer

Item Location : NA

Make / Model

Serial / I.D. No.

Inst. Range

Least Count

Unit

Item Condition

NA / NA

NA / NA

0 min to 60 min

1

min

Working

Environmental Condition

25.75 ± 0.5 °C

52 ± 3% RH

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard

Make / Model

Certificate No.

Calibration Agency

Valid up to

Time Interval Meter

ZEAL / ZMTIM

AACPL/21357F

AA Calibration Lab Pvt. Ltd.

20-10-2023

Calibration Procedure

Lab Procedure : MTE-2/ET/2.19

STD Procedure : NIST 960-12

RESULTS: In-Lab Calibration

All Readings are in min, Specified Otherwise.

UUC READING

CORRECTED VALUE OF
STANDARD (AVERAGE) (H:M:S)

ERROR

UNCERTAINTY[^]

5

00:04:59.826

0.174

± 0.01

10

00:09:59.630

0.370

± 0.01

15

00:14:59.868

0.132

± 0.01

30

00:29:59.975

0.025

± 0.01

[^]The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

For and on behalf of
Multitech Technofab Enterprises

Calibrated By MANAVI SINHA
Deputy Quality Manager

Laboratory Head



Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area
Delhi - 110092

Phone No. : 09810943334 , 011-41764006, Email:
multitech.technofab@gmail.com, calibration@multitechtechnofab.com , Web. :
www.multitechtechnofab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/20052

ULR :
CC360323000002331F

Calibration Issued on: 27-09-2023

Note

1. This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
2. The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
3. Total liability of our organization is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

For and on behalf of
Multitech Technofab Enterprises

Calibrated By MANAVI SINHA
Deputy Quality Manager

Laboratory Head