



# CALIBRATION CERTIFICATE

Page No: 1 of 1

Certificate No: SBS/CL/23/19952	SRF No.	SRF/23/00842-0004
Customer Name & Address  GOVERNMENT PRIMARY HEALTH CENTRE, PERUMBAKKAM-631551, KANCHIPURAM DISTRICT.	SRF Date	09-10-2023
	Date of Receipt	07-10-2023
	Date of Calibration	09-10-2023
	Due Date for Calibration	08-10-2024
	Issue Date	10-10-2023

Details of Unit Under Calibration		Make	THERMO SCIENTIFIC
Description	MICRO PIPETTE	Model	FINNPIPETTE F3
Range	10-100µl	Material	PVC
Resolution	0.2µl	Operating Range	10-100µl
Serial Number	NA	Condition of UUC	Good
ID Number	NA	Instrument Location	LABORATORY
Cal. At	Mechanical Lab		

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08


SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP/22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
SI. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47	
		9.98	9.96	9.97	9.98	9.95					
2	50.0	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47	
		49.95	49.96	49.97	49.95	49.95					
3	100.0	99.98	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47	
		99.95	99.96	99.94	99.95	99.96					

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized Water conforming Grade3 as specified in ISO 3696.

Calibrated By,  
  
 (Calibration Engineer)  
 M.RAGUL



Authorized by:  
  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

**SUNSHINE BIOMEDICAL SOLUTIONS**  
 No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.