

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Primary Health Center, #Keelachekkarakudi-628104.	Certificate No:	AMS/LAB/289/2522
		Issuing Date:	05/11/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	KPHC/LAB/CEN/001
Make:	Dalal	Department:	Laboratory
Model:	Digital	Calibrated On:	03/11/2022
Sl. No:	---	Calibration Due Suggested:	02/11/2023
Type:	Digital – Manual Control	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 4000 / LC – 1; Timer 10 to 60Mins. In 10Mins. interval		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

### Test Results: Rotation Per Minute

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1150	0
2	Position 2	2180	0
3	Position 3	2540	0
4	Position 4	3150	0
5	Position 5	3850	0

Allowable Tolerance: Set by Customer

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:  
(Calibration Engineer)

aceagm@gmail.com



amshealthcaremdu@gmail.com

Approved by:  
(Technical Manager)

94434 74610, 93455 56148