

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

| | | | |
|--------------------------|---|-----------------|------------------|
| Customer Name & Address: | M/S. Primary Health Center, #Keelachekkarakudi-628104. | Certificate No: | AMS/LAB/289/2526 |
| | | Issuing Date: | 05/11/2022 |
| | | Page No: | 1 of 1 |

Details of Device under Calibration (DUC):

| | | | |
|----------------------|--------------|----------------------------|------------------|
| DUC: | Micropipette | Id. No: | KPHC/LAB/VMP/003 |
| Make: | P Fact | Department: | Laboratory |
| Model: | --- | Calibrated On: | 03/11/2022 |
| Sl. No: | 230531 | Calibration Due Suggested: | 02/11/2023 |
| Type: | Variable | DUC Condition on Receipt: | OK |
| Range & Least Count: | 100 – 1000µl | | |

Traceability Details:

| Sl.No. | Test Equipment | Make | Certificate Number | Valid Up to |
|--------|---------------------------|-------------|----------------------|-------------|
| 1 | Electronic Weighing Scale | EWC/SF400C | PC/01/22/ATL/0432-02 | 30/01/2023 |
| 2 | Thermo Hygrometer | Beetech/088 | PC/01/22/ATL/0265-01 | 30/01/2023 |

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results Volume µl:

| SL. No. | DUC Reading (µl) | Standard Reading (µl) | | Correction (µl) |
|---------|------------------|-----------------------|---------|-----------------|
| | | Trial 1 | Trial 2 | |
| 1 | 100.0 | 110.00 | 110.00 | +10.00 |
| 2 | 200.0 | 210.00 | 210.00 | +10.00 |
| 3 | 500.0 | 500.00 | 500.00 | +00.00 |
| 4 | 800.0 | 800.00 | 800.00 | +00.00 |
| 5 | 1000.0 | 990.00 | 990.00 | -10.00 |

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

Approved by:

(Technical Manager)