

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Virachilal
 Lab No 12525
 Month August
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 17/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	34	172.03	5.51	9.48	162 mg/dL	-1.06	3.25
2	CREATININE III	Jaffes Kinetle - Alkaline Picrate III	Biosystems	35	1.76	6.47	0.11	1.9 mg/dL	1.23	0.04
3	CHOLESTROL III	CHOD-PAP III	Biosystems	33	114.98	3.90	4.49	114 mg/dL	-0.22	1.56

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

P. Shanil
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 VIRACHILAI-622 412
 PUDUKKOTTAI DISTRICT