

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Primary Health Center, #Keelachekkarakudi-628104.	Certificate No:	AMS/LAB/289/2527
		Issuing Date:	05/11/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	KPHC/LAB/VMP/002
Make:	Diagnopette	Department:	Laboratory
Model:	---	Calibrated On:	03/11/2022
Sl. No:	181250425	Calibration Due Suggested:	02/11/2023
Type:	Variable	DUC Condition on Receipt:	OK
Range & Least Count:	5-50 μ l		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C \pm 2 Room Humidity: 30-60%

(At the time of Calibration)

Test Results Volume μ l:

SL. No.	DUC Reading (μ l)	Standard Reading (μ l)		Correction (μ l)
		Trial 1	Trial 2	
1	5.0	04.50	04.50	-0.50
2	10.0	09.00	09.00	-1.00
3	20.0	18.00	18.00	-2.00
4	30.0	28.00	28.00	-2.00
5	50.0	50.00	50.00	-0.00

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)