

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Virachilal
 Lab No 12525
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Sl No.	Analyte	Method Reagent Name	Analyzer Name	No of Participants	UV	Participants CV	CV	CV	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	36	280.51	4.07	11.42	286 mg/dL	0.48	3.81
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	39	2.67	6.50	0.17	2.6 mg/dL	-0.40	0.06
3	CHOLESTROL III	CHOD-PAP III	Biosystems	37	113.79	4.23	4.81	110 mg/dL	0.46	1.58

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Today Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign out

R. Shanmugam
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 VIRACHILAL-622 412
 PUDUKKOTTAI DISTRICT