



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. PHC, Virachilai
Constituent Group Chemistry III
PT Item Lyophilized human serum based

Lab No
Date of Result Entered :
Date of Report Published :

12525
20/05/2023
05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85	188 mg/dL	0.83	3.43
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16	1.7 mg/dL	-0.50	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47	116 mg/dL	-0.07	2.49

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Virachilai
Virachilai
TAMILNADU

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

R. Sharmila
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
VIRACHILAI - 622 412
PUDUKKOTTAI DISTRICT