

VIEW LAB MONTHLY ALL RESULT SUMMARY

| Lab Name | Govt. PHC, Pulvayal | | | | | | | | | |
|---------------------------------------|---------------------|---------------------------------------|----------------------------------------------|--------------------|--------|-----------------|-------|------------|-------|------|
| Lab No | 12506 | Back to Search Page | | | | | | | | |
| Month | September | | | | | | | | | |
| Year | 2023 | <input type="button" value="Print"/> | | | | | | | | |
| Constituent Group | Chemistry III | | | | | | | | | |
| Date of Result Entered : 20/09/2023 | | | | | | | | | | |
| Date of Report Published : 05/10/2023 | | | | | | | | | | |
| Sl.No | Constituent Name | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants CV | SD | Your Value | SDI | U |
| 1 | GLUCOSE III | GOD-POD III | Biosystems | 36 | 312.41 | 4.03 | 12.59 | | | 4.20 |
| | | | Transasia / Erba | 18 | 307.20 | 6.79 | 20.85 | | | 9.83 |
| | | | Robonik (Semi automatic / Fully Automatic) | 1870 | 310.86 | 4.82 | 14.98 | 304 mg/dL | -0.46 | 0.69 |
| | | | Any Analyser (Automation / Semi Automation) | 68 | 312.35 | 4.82 | 15.05 | | | 3.65 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Biosystems | 38 | 1.30 | 7.23 | 0.09 | | | 0.03 |
| | | | Transasia / Erba | 21 | 1.26 | 9.52 | 0.12 | | | 0.05 |
| | | | Robonik (Semi automatic / Fully Automatic) | 1852 | 1.29 | 8.14 | 0.10 | 1.3 mg/dL | 0.10 | 0.00 |
| | | | Any Analyser (Automation / Semi Automation) | 69 | 1.31 | 8.62 | 0.11 | | | 0.03 |
| 3 | CHOLESTROL III | CHOD-PAP III | Biosystems | 35 | 113.55 | 3.41 | 3.87 | | | 1.31 |
| | | | Transasia / Erba | 19 | 113.64 | 2.93 | 3.33 | | | 1.53 |
| | | | Robonik (Semi automatic / Fully Automatic) | 1783 | 113.62 | 3.80 | 4.32 | 105 mg/dL | -2.00 | 0.20 |
| | | | Any Analyser (Automation / Semi Automation) | 151 | 113.39 | 4.43 | 5.02 | | | 0.82 |
| SDI Range | | | Interpretation | | | | | | | |
| Within -1.00 to +1.00 | | | Excellent. | | | | | | | |
| Within ±1.01 to ±2.00 | | | Good. | | | | | | | |
| Within ±2.01 to ±2.99 | | | Accept with caution. Warning Signal. | | | | | | | |
| Beyond ±3.0 | | | Unacceptable performance. Action Signal. | | | | | | | |

Links

| | | | |
|-----------------------|------------------------|---------------------------|---------------------------------|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |