



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Mokuveri
Constituent Group: Chemistry III
PT item: Lyophilized human serum based

Lab No: 9256
Date of Result Entered: 13/04/2023
Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121.5 mg/dL	0.28	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.52 mg/dL	2.06	0.01
	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	113.7 mg/dL	0.36	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Mokuveri
THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, MUKUPERI, MUKKUPERI-PO, THIRUCHANDUR TK, THOOTHUKUDI
Mokuveri
MILNADU 628616

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

Smy.V
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
MUKKUPERI-PO, THIRUCHANDUR TK,
THOOTHUKUDI, DIST. THIRUVARUR



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MARCH 2023



Lab Name: Govt. PHC, Mokuiperi
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 9266
 Date of Result Entered: 16/03/2023
 Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Mean Value	SD
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic	1878	91.48	7.46	6.82	81.7 mg/dL	-4.43
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic	1851	1.33	9.08	0.12	1.23 mg/dL	-0.83
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic	1761	110.93	5.23	5.80	110 mg/dL	-8.16

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Mokuiperi
 THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, MUKUPERI, MUKKUPERI-PO, THIRUCHANDUR TK, THOOTHUKUDI
 Mokuiperi
 TAMILNADU628616

Smy.V
 MEDICAL OFFICER,
 GOVT. PRIMARY HEALTH CENTRE
 MUKKUPERI - 628 616
 THOOTHUKUDI DT.
Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Mokuiperi
 Lab No 9266
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 18/06/2023

Date of Report Published : 04/07/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	303.4 mg/dL	1.13	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.62 mg/dL	0.29	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	110.7 mg/dL	-0.60	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Smit V
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 MOKUPIERI - 628 616
 THOOTHUKUDI DT.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

https://home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/ViewResult2_new.aspx?resDetId=266&noSamples=1&month=6&year=2023&groupId=9&yearGroup=110

[View Lab Monthly Summary](#)

[Sign-out](#)



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. PHC, Mokuveri Lab No 9266
 Constituent Group Chemistry III Date of Result Entered : 15/05/2023
 PT item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178.6 mg/dL	-0.24	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.86 mg/dL	0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	112.1 mg/dL	-0.43	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Mokuveri

THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, MUKUPERI, MUKKUPERI-PO, THIRUCHANDUR TK, THOOTHUKUDI Mokuveri

TAMILNADU 628616

TAM

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

Smf V
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 MOOKUPERI - 628 616
 THOOTHUKUDI DT.



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Mookuperi
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No: 9256
 Date of Result Entered: 18/07/2023
 Date of Report Published: 07/08/2023

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	SD	Participants	Year Value	SE
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic	1876	5.55	6.07	109.35	109 mg/dL	-0.06
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic	1850	6.00	0.35	5.83	5 mg/dL	-2.37
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic	1776	4.12	4.66	113.00	103 mg/dL	-2.15

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Mookuperi
 THE MEDICAL OFFICER, GOVT.PRIMARY HEALTH CENTRE, MUKUPERI, MUKKUPERI, THIRUCHANDUR TK, THOOTHUKUDI
 Mookuperi
 TAMILNADU 628616

Smt. V
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
MOOKUPERI - 628 616
THOOTHUKUDI DT.

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****