



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. PHC, Mullukurichi      Lab No: 12439  
 Constituent Group: Chemistry III      Date of Result Entered: 20/06/2023  
 PT item: Lyophilized human serum based      Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	Result	Reference Range	CV (%)	CV (%)	CV (%)	CV (%)	CV (%)
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	305.3 mg/dL	1.26	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.72 mg/dL	0.88	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	116.2 mg/dL	0.49	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. PHC, Mullukurichi  
 CLINICAL LABORATORY  
 Mullukurichi  
 TAMILNADU637406

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
**MEDICAL OFFICER,**  
 Govt. Primary Health Centre,  
 MULLUKURICHI-636 142,  
 Rasipuram(Tk), Namakkal(Dt)