

8/11/23, 4:03 PM

External Quality Assurance Scheme - Print Monthly Summary



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



|                   |                               |                            |            |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name          | Govt. PHC, Thittacheri        | Lab No                     | 12421      |
| Constituent Group | Chemistry III                 | Date of Result Entered :   | 20/07/2023 |
| PT item           | Lyophilized human serum based | Date of Report Published : | 07/08/2023 |

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |            |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1876               | 109.35 | 5.55         | 6.07 | 108 mg/dL  | -0.22 | 0.28 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1850               | 5.83   | 6.00         | 0.35 | 5.7 mg/dL  | -0.37 | 0.02 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1776               | 113.00 | 4.12         | 4.66 | 118 mg/dL  | 1.07  | 0.22 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

## LAB ADDRESS :

Govt. PHC, Thittacheri  
CLINICAL LABORATORY  
Thittacheri  
TAMILNADU

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Handwritten signature]*  
17/08/23