



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL CHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. PHC, Thittacheri
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based

Lab No: 12421
 Date of Result Entered: 19/04/2023
 Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	182	119.45	6.04	7.21	103 mg/dL	-2.28	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	182	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	175	111.61	5.19	5.79	120 mg/dL	1.45	0.28

come

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent performance
Within ±1.01 to ±2.00	Good performance
Within ±2.01 to ±2.99	Acceptable caution, Watch for Signal.
Beyond ±3.0	Unsatisfactory performance, Action Signal.

LAB ADDRESS :
 Govt. PHC, Thittacheri
 CLINICAL LABORATORY
 Thittacheri
 TAMILNADU

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

come the Glucose Value

Homogeneity and stability of the test is passed.
 Data is complete and report is generated.
 CMC EQAS approved sub-centre for participants
 for the month of April 2023

Signature