



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 14.09.2023
	Calibration Valid: 13.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVT.PRIMARY HEALTH CENTRE, ARNOOTHUMALAI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	CENTRIFUGE
i	Manufactured by BD INSTRUMENTATION
ii	Model BDI-152
iii	Serial No PHC/ATI/CF/001
iv	Location LABORATORY

CALIBRATION WORKSHEET

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM
1	1000	1002	2
2	1200	1200	0
3	1500	1501	1
4	2000	2000	0

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT-100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non-contact) measured has been inspected in accordance with our inspection procedures to meet manufactures specifications.

Remarks

This certificate may not be reproduced other than in full Except prior permission of laboratory.

Verified by : 
