




**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 EQAS YEARLY SUMMARY - 2022



Lab Name: Govt. PHC, Arunuthumalai  
 Lab No: 12576  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Date of Report Published: 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022  
 Group II : Laboratories who reported 6 to 9 months during 2022  
 Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.71	A
2	CREATININE III	Group II	99	1.42	B
3	CHOLESTROL III	Group I	1923	1.41	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond 3.00	D	Unacceptable

LAB ADDRESS  
 Govt. PHC, Arunuthumalai  
 Arunuthumalai  
 TAMILNADU636139

Coordinator Contact Details:  
 Email: [clinqc@cmcvallure.ac.in](mailto:clinqc@cmcvallure.ac.in)  
 Contact number: 0416-2283102

*E. Leela Devi*

**DR. R. LEEELA DEVI, M.B.B.S.,**  
 SENIOR ASSISTANT SURGEON,  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUNUTHUMALAI, SALEM (Dt).

Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christodoss*  
**Dr. Pamela Christodoss**  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Arunthumalai

Lab No: 12576

Constituent Group: Chemistry III

Date of Result Entered: 20/04/2023

PT Item: Lyophilized human serum based

Date of Report Published: 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	132 mg/dL	1.74	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	126 mg/dL	2.49	0.28
<b>SDI Range Interpretation</b>										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
 Govt. PHC, Arunthumalai  
 Arunthumalai  
 Arunthumalai  
 TAMILNADU636139

**Coordinator Contact Details:**  
 Email: cllngc@cmcvellore.ac.in  
 Contact Number: 0416-2283102  
*P. R. S. S. S.*

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EOAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EOAS reports is confidential  
 CMC EOAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. PHC, Arunthumalai      Lab No: 12576  
 Constituent Group: Chemistry III      Date of Result Entered: 22/05/2023  
 PT Item: Lyophilized human serum based      Date of Report Published: 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SO	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	180 mg/dL	-0.09	0.44
2	CREATININE III	Jaffes Kinetiic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.5 mg/dL	-1.97	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
<b>SDI Range</b>										
				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**

Govt. PHC, Arunthumalai  
 arunthumalai  
 Arunthumalai  
 TAMILNADU636139

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102  
*R. Sulekha*

**DR. R. LEEBA DEVI, M.B.B.S.,**  
 SENIOR ASSISTANT SURGEON,  
 REG. NO: 85370  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUNTHUMALAI, SALEM (DI).

*Pamela Christudassi*  
 Dr. Pamela Christudassi  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*



**CHRISTIAN MEDICAL COLLEGE**

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JUNE 2023

PC-1024

Lab Name: Govt. PHC, Arunthumalai  
 Lab No: 12576  
 Constituent Group: Chemistry III  
 Date of Result Entered: 20/06/2023  
 PT item: Lyophilized human serum based  
 Date of Report Published: 04/07/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	297 mg/dL	0.70	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.2 mg/dL	-2.16	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	129 mg/dL	1.24	0.24
<b>SDI Range</b>										
				<b>Within -1.00 to +1.00</b>						
				Excellent.						
				<b>Within ±1.01 to ±2.00</b>						
				Good.						
				<b>Within ±2.01 to ±2.99</b>						
				Accept with caution. Warning Signal.						
				<b>Beyond ±3.0</b>						
				Unacceptable performance. Action Signal.						
<b>Interpretation</b>										

**LAB ADDRESS :**  
 Govt. PHC, Arunthumalai

arunthythumalai  
 Arunthumalai  
 TAMILNADU636139

**Coordinator Contact Details:**  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0415-2283102

**Dr. R. LEELA DEVI, M.B.S.**  
 SENIOR ASSISTANT SURGEON,  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUMTHUMALAI, SALEM (DI).

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. PHC, Arunthumalai  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based

Lab No: 12576  
 Date of Result Entered: 20/07/2023  
 Date of Report Published: 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	118 mg/dL	1.43	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.1 mg/dL	0.77	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22
SDI Range										
Within -1.00 to +1.00				Excellent						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						
Interpretation										

LAB ADDRESS :  
 Govt. PHC, Arunthumalai  
 arunthumalai  
 Arunthumalai  
 TAMILNADU636139

Coordinator Contact Details:  
 Email:clinqgc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

Dr. R. LEEBA DEVI, M.B.B.S.,  
 SENIOR ASSISTANT SURGEON,  
 Reg. No: 85370  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUNTHUMALAI, SALEM (Dt).

*Pamela Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

**Lab Name** Govt. PHC, Arunuthumalai  
**Constituent Group** Chemistry III  
**PT Item** Lyophilized human serum based

**Lab No** 12576  
**Date of Result Entered :** 19/08/2023  
**Date of Report Published :** 01/09/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Year Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonk ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	168 mg/dL	-0.84	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonk ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	106 mg/dL	-1.68	0.22
<b>SDI Range</b>										
				<b>Within -1.00 to +1.00</b>						
				Excellent.						
				<b>Within ±1.01 to ±2.00</b>						
				Good.						
				<b>Within ±2.01 to ±2.99</b>						
				Accept with caution. Warning Signal.						
				<b>Beyond ±3.0</b>						
				Unacceptable performance. Action Signal.						

**LAB ADDRESS :**  
 Govt. PHC, Arunuthumalai  
 arunuthumalai  
 Arunuthumalai  
 TAMILNADU636139

**Coordinator Contact Details:**  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

**Dr. R. A. E. S. M. B. S.**  
 SENIOR ASSISTANT SURGEON,  
 Reg. No: 85370  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUNUTHUMALAI, SALEM (Dt).

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*Ravula Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Yelore



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. PHC, Arunthumalai  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based

Lab No: 12576  
 Date of Result Entered: 20/09/2023  
 Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U		
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	310 mg/dL	-0.06	0.69		
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00		
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	111 mg/dL	-0.61	0.20		
<b>SDI Range Interpretation</b>												
				Within -1.00 to +1.00							Excellent.	
				Within ±1.01 to ±2.00							Good.	
				Within ±2.01 to ±2.99							Accept with caution. Warning Signal.	
				Beyond ±3.0							Unacceptable performance. Action Signal.	

**LAB ADDRESS :**  
 Govt. PHC, Arunthumalai  
 arunuyhumalai  
 Arunthumalai  
 TAMILNADU636139

**Coordinator Contact Details:**  
 Email: cllnqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102  
**Dr. R. J. DEVI, M.B.B.S.,**  
 SENIOR ASSISTANT SURGEON,  
 Reg. No: 85370  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 ARUNTHUMALAI, SALEM (Dt).

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*