

## VIEW LAB MONTHLY ALL RESULT SUMMARY

Lab Name	Govt. PHC, Pulvayal									
Lab No	12506	<a href="#">Back to Search Page</a>								
Month	May									
Year	2023	<input type="button" value="Print"/>								
Constituent Group	Chemistry III									
Date of Result Entered : 22/05/2023										
Date of Report Published : 05/06/2023										
Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Biosystems	33	179.81	5.48	9.85			3.43
			Transasia / Erba	19	177.53	5.62	9.98			4.58
			Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	185 mg/dL	0.44	0.44
			Any Analyser (Automation / Semi Automation )	69	181.20	5.81	10.52			2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	38	1.78	8.91	0.16			0.05
			Transasia / Erba	22	1.85	7.73	0.14			0.06
			Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.6 mg/dL	-1.21	0.01
			Any Analyser (Automation / Semi Automation )	68	1.75	7.05	0.12			0.03
3	CHOLESTROL III	CHOD-PAP III	Biosystems	36	116.50	6.42	7.47			2.49
			Transasia / Erba	20	115.73	3.06	3.54			1.58
			Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	108 mg/dL	-1.13	0.27
			Any Analyser (Automation / Semi Automation )	154	114.84	5.08	5.83			0.94
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

**Links**

<b>Modify Address</b>	<b>Modify Methods</b>	<b>Submit Lab Results</b>	<b>View Lab Monthly Summary</b>
<b>Yearly Summary</b>	<b>Change Password</b>	<b>Home</b>	<b>Sign-out</b>