



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

External Quality Assurance Scheme - Print Monthly Summary

MDY-23

Lab Name Govt. Block PHC, Rishivandiyam Lab No 7289
 Constituent Group Chemistry III Date of Result Entered : 20/05/2023
 PT Item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	178 mg/dL	-0.31	0.44
2	CREATININE III	Jafes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27
SDI Range										
				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. Block PHC, Rishivandiyam
 RISHIVANDIYAM PHC
 Rishivandiyam
 TAMILNADU606205

Coordinator Contact Details:
 Email:clinq@cmcvellore.ac.in
 Contact Number: 0416-2233102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

6/28/23