

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. UPHC, Ganesh Naghar
 Lab No: 10697
 Month: March
 Year: 2023
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Transasia / Erba	18	92.24	11.77	10.85	114.7 mg/dL	2.07	5.12
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	55.01	12.13	6.67	45.19 mg/dL	-1.47	3.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	16	1.36	9.29	0.13	1.75 mg/dL	3	0.06
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	111.05	6.05	6.72	109.1 mg/dL	-0.29	3.08
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Digitally Signed
 Digitally signed by Ganesh Naghar
 Date: 2023.03.20 10:15:30 +05'30'

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Ganesh Naghar	Details About Robust Analysis
Lab No	10697	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	118.42	8.48	10.04	106.5 mg/dL	-1.19	4.61
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	41.88	8.81	3.69	38.52 mg/dL	-0.91	2.13
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	19	2.14	6.03	0.13	2.08 mg/dL	-0.47	0.06
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	111.15	4.25	4.72	110.3 mg/dL	-0.18	2.11

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Ganesh Naghar	Details About Robust Analysis
Lab No	10697	Detail About Monthly Summary
Month	May	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 17/05/2023


Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	177.53	5.62	9.98	167.1 mg/dL	-1.05	4.58
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	11	25.79	11.91	3.07	38.62 mg/dL	4.18	1.85
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	22	1.85	7.73	0.14	1.93 mg/dL	0.56	0.06
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	115.73	3.06	3.54	115 mg/dL	-0.21	1.58

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 Digitally signed by Govt. UPHC, Ganesh Naghar
 Date: 2023.06.05 10:00:00 +05'30'



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. UPHC, Ganesh Naghar Lab No 10697
Constituent Group Chemistry III Date of Result Entered : 20/06/2023
PT item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	22	283.77	5.80	16.44	272 mg/dL	-0.72	7.01
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	12	31.96	9.97	3.19	28 mg/dL	-1.24	1.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	23	2.50	9.65	0.24	2.4 mg/dL	-0.41	0.10
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	22	114.61	5.43	6.22	117 mg/dL	0.38	2.65

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Ganesh Naghar
DR. Arthy, GANESH NAGAR
Ganesh Naghar
TAMILNADU628008

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****

(Handwritten signature)
சென்னை
கி.மீ.சி.ம.க. விலை
கி.மீ.சி.ம.க. விலை

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. UPHC, Ganesh Naghar
 Lab No 10697
 Month July
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Transasia / Erba	24	109.83	6.83	7.50	110 mg/dL	0.02	3.06
2	UREA III	Urease UV / GLDH III	Any Analyser (Automation / Semi Automation)	11	50.80	9.44	4.79	54 mg/dL	0.67	2.89
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	18	5.79	3.00	0.17	5.8 mg/dL	0.06	0.08
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	112.64	4.38	4.93	113 mg/dL	0.07	2.26
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

[Yearly Summary](#)

[Change Password](#)

[Home](#)

[Sign-out](#)

Handwritten signature and notes in blue ink:
 Dr. Ganesh Naghar
 Govt. UPHC, Ganesh Naghar
 July 2023



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name: Govt. UPHC, Ganesh Naghar

Lab No: 10697

Constituent Group: Chemistry III

Date of Result Entered: 19/09/2023

PT item: Lyophilized human serum based

Date of Report Published: 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Transasia / Erba	18	307.20	6.79	20.85	270 mg/dL	-1.78	9.83
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	187	35.19	12.73	4.48	40 mg/dL	1.07	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.26	9.52	0.12	1.2 mg/dL	-0.50	0.05
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	113.64	2.93	3.33	114 mg/dL	0.11	1.53
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

மேன்மை சீரமைப்பு
கொடுக்கப்பட்டது

மேன்மை சீரமைப்பு
கொடுக்கப்பட்டது

[Handwritten signature]

LAB ADDRESS :
Govt. UPHC, Ganesh Naghar
DR. Arthy, GANESH NAGAR
Ganesh Naghar
TAMILNADU628008

Pamela Christodoss

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Coordinator Contact Details:
Email: clinqcc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.