



CHRISTIAN MEDICAL COLLEGE

Aug-23



PC-1024

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023

Lab Name: Govt. Block PHC, Rishivandiyam
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based
 Lab No: 7289
 Date of Result Entered: 19/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	178 mg/dL	0.27	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2 mg/dL	2.02	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22
SDI Range					Interpretation					
Within -1.00 to +1.00					Excellent.					
Within ±1.01 to ±2.00					Good.					
Within ±2.01 to ±2.99					Accept with caution. Warning Signal.					
Beyond ±3.0					Unacceptable performance. Action Signal.					

LAB ADDRESS :
 Govt. Block PHC, Rishivandiyam
 RISHIVANDIYAM PHC
 Rishivandiyam
 TAMILNADU606205

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****