



# Vanavil Calibrations

Make It Simple, But Significant!

## CALIBRATION CERTIFICATE

Certificate No: VCPL/CC/2023/014-004

Page no: 1 of 2

Customer Name & Address:  M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE, GANESH NAGAR, THOOTHUKUDI - 628 008.	SRF NO	SRF /23-24/014-004 & 07/10/2023
	Date of Calibration:	7-Oct-2023
	Date of issue:	8-Oct-2023
	Date of Receipt:	7-Oct-2023
	Next cal Date:	6-Oct-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	FINNPIPETTE F3
Range:	(100 TO 1000) $\mu$ l	Model/Type:	-
Resolution:	1 $\mu$ l	Condition of DUC	GOOD
Serial Number	RW13768	Location:	LABORATORY
ID number	-	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.5 $^{\circ}$ C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	52 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION

S.NO	Duc Readings ( $\mu$ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100	99.89	0.11
2	500	499.85	0.15
3	1000	999.79	0.21

### Remarks:


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
C.AKASH

(CALIBRATION ENGINEER)



Authorised by: 

M. KANNAN,  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com