



# Vanavil Calibrations

Make It Simple, But Significant !

## CALIBRATION CERTIFICATE

Certificate No: VCPL/CC/23-24/014-002

Page : 1 of 1

Customer Name & Address:  M/S GOVERNMENT URBAN PRIMARY HEALTH CENTRE, GANESH NAGAR, THOOTHUKUDI - 628 008.	SRF NO & DATE	SRF/23-24/017-002 & 07/10/2023
	Date of Calibration:	7-Oct-2023
	Date of issue:	8-Oct-2023
	Date of Receipt:	7-Oct-2023
	Next cal Date:	06-10-2024

### Details of Device Under Calibration

Description	CENTRIFUGE	Make:	-
Range	3000 RPM	Model/Type:	-
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	20111150	Instrument Location:	LABORATORY

### Enviromental Conditions

Temperature	23.4 °C	Reference Standard No	SANAS TR4501
Humidity	52 % Rh	Calibration Procedure No	-

### Method Used For Calibration

### Details of Reference Standard Equipment For Calibration

S.No	Description	ID No / SL No	Certificate No	Validity
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005 SYSTEMS/ HTM 560	4-Dec-23


### Calibration Result

S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	3000	2999	1	12


### Remarks:


- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
C.AKASH

(CALIBRATION ENGINEER)



Authorised by: 

M. KANNAN,  
Technical Manager  
Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com