

## ROOTS INDUSTRIES INDIA LIMITED

# ROOTS METROLOGY & TESTING LABORATORY

R.K.G Industrial Estate, Ganapathy, Coimbatore-641 006,India Phone: 9585840192/9003432356





CERTIFICATE NO CC-2752

## CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer

The Medical officer

Government Primary Health Center, Embal-

Pudhukottai( dist)

Certificate no

: Rmtl/01/319107848/15

Date of calibration

:03/08/2023

Suggested calibration due

: 02/08/2024

Customer po .reference.& date:

DESCRIPTION OF DUC

Name Of The Duc:

Pipette

**THERMOSCIENTIFIC** 

Receiving Condition Of Duc

:Good

Make : SI.No :Rw10851

Model :Finnpipette.

**Environmental Condition** 

Ref. Document : As Per Manufactures Specification :25.3 Deg & 54.3% Rh

Uncertainty Of Measurement

:As Per Report

Range/Size

:10micro litter

Location :lab

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

### **DETAILS OF MASTER USED**

| Standard Instrument Used                                     | Make/ Model          | ld No   | Traceable To/Certificate No | Validity |
|--|----------------------|---------|-----------------------------|----------|
| Multi Product Calibrator                                     | Fluke                | Rml-262 | Wavecal, Blr/Cs-18/01-01    | 1-6-2024 |
| Micro balance  | mettle toledo/XP26PC | QP-01   | B115129039/CC-2523          | 1-6-2024 |
| Universal Measuring Instrument With Sensor For Environmental | ALMEMO/2590-4AS      | QP-10   | H19070544/CC-2052           | 1-6-2024 |
| Parameter  |                      |         |                             |          |

# **CALIBRATION FIELD: WEIGHT Calibration**

| Sl.no  | Range          | Master readings micro litter | Calibrated values micro litter | Error observed micro litter | Uncertainty |
|--------|----------------|------------------------------|--------------------------------|-----------------------------|-------------|
| 1      | 10micro litter | 10                           | 9.9                            | 0.1                         | 0.055       |
| 2      | 10micro litter | 10                           | 9.8                            | 0.1                         | 0.055       |
| 3      | 10micro litter | 10                           | 9.7                            | 0.1                         | 0.055       |
| л<br>Л | 10micro litter | 10                           | 9.5                            | 0.1                         | 0.055       |
| 5      | 10micro litter | 10                           | 9.8                            | 0.1                         | 0.055       |

uncertainty estimated for a confidence level of 95% at the k factor=2

the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

DUC -device under calibration

Calibration label /status : Calibration result pass & fail are indicated

CALIBRATION ENGINEER

**AUTHORIZED SIGN** 

This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell:9585840192