



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 30.09.2023
	Calibration Valid: 29.03.2024
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HESLTH CENTRE, SOLAKKADU, NAMAKKAL.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
<b>i</b> Manufactured by	REMI
<b>ii</b> Model	R8C
<b>iii</b> Serial No	ZEAN 00721
<b>iv</b> Location	LABORATORY

### CALIBRATION WORKSHEET

#### Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navalindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com

REDMI NOTE 6 PRO  
MI DUAL CAMERA