



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/12498</b>	<b>MEDICAL DEVICES</b>	Page No:1 of 1
Issue Date	30-09-2023	
SRF No & Date	SRF/23/00553-0002 & 28-09-2023	
Receipt Date	28-09-2023	
Calibration Date	28-09-2023	
Calibration Due	27-09-2024	

**Customer Name & Address**  
GOVERNMENT PRIMARY HEALTH CENTRE,  
KEELAPOONGUDI-630552.

Details of Device Under Calibration (DUC)			
Description	ELECTRICAL SAFETY(CELL COUNTER)	Make & Model	SYSMEX & XP-300
Range	MULTI	Sr. No	B7810
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resisitance</b>	<b>Measured values in MΩ</b>	<b>Uncertainty in % ( ± )</b>
	>20MΩ	98	13.92
2	<b>Earth Leakage</b>	<b>Measured values in μA</b>	<b>Uncertainty in % ( ± )</b>
	<5000μAfor B,BF,CF	204	9.1
3	<b>Enclosure Leakage</b>	<b>Measured values in μA</b>	<b>Uncertainty in % ( ± )</b>
	<500μAfor B,BF,CF	214	8.9

- REMARKS**
- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
  - 2.The user should determine the suitability of the instrument for its intended use.
  - 3.The recalibration interval should be determined on the user requirement.
  - 4.The results stated in this certificate relate only to the item calibrated.
  - 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
  6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
P.PRASANNA



Authorised Signatory

  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.