

		<u>CALIB</u>	RATION	CERTII	FICATE			
CERTIFICATE NO:	SBS/C	L/23/12500		М	ECHANICAL	Pa	ge No:1 of 1	
Issue Date			30-09	-2023				
SRF No & Date			SRF/23/00553-0004 & 28-09-2023					
Receipt Date				28-09-2023				
Calibration Date			28-09	28-09-2023				
Calibration Due			27-09	27-09-2024				
Customer Name & Ad	dress							
GOVERNMENT PRI	MARY HEAL	TH CENTRE,						
KEELAPOONGUDI-6	30552.	,						
		Detai	ls of Device Und	der Calibratio	n (DUC)			
Description :	CENTE	RIFUGE			:	REMI & R-8C		
Range :	5250	RPM	Sr. No		:	ZHBN-07180		
Least Count :	10	RPM	Identif	ication No	:	NA		
DUC Condition :	Satisfa	ctory	Locati	on	4	LAB		
	Env	ronmental Co	nditions & Stan		g Procedure			
Environmental Details Temperatur						57% Rh		
Calibration Procedure No SBS/CP/M								
		•	Reference Star			15.15.112		
S.No Description	Description		Make/ SI No:		Certificate No		Validity	
Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-2024	
CALIBRATION RESU	ILTS							

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.7	0.3	4.2
2	2000	1999.7	0.3	4.2
3	3000	2999.6	0.4	10
REMAR	Ke		0.1	4.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By (Calibration Engineer) P.PRASANNA



C.SIVABALAN

Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS