



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. Block PHC, Solakkadu  
Constituent Group: Chemistry III  
PT item: Lyophilized human serum based

Lab No: 7446  
Date of Result Entered: 19/06/2023  
Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	309.49 mg/dL	1.54	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.51 mg/dL	-0.35	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	117.43 mg/dL	0.73	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within $\pm 1.01$ to $\pm 2.00$				Good.						
Within $\pm 2.01$ to $\pm 2.99$				Accept with caution. Warning Signal.						
Beyond $\pm 3.0$				Unacceptable performance. Action Signal.						

LAB ADDRESS :  
Govt. Block PHC, Solakkadu  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT.  
Solakkadu  
TAMILNADU637411

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Paula Christodoss*  
Dr. Pamela Christodoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*[Signature]*  
MEDICAL OFFICER  
Government Primary Health Centre  
Solakkadu, Koli Hills,  
Namakkal - 637411.

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*