



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO:</b> SBS/CL/23/13924	<b>MEDICAL DEVICES</b>	<b>Page No:</b> 1 of 1
Issue Date	10-10-2023	
SRF No & Date	SRF/23/00840-0001 & 09-10-2023	
Receipt Date	09-10-2023	
Calibration Date	09-10-2023	
Calibration Due	08-10-2024	

**Customer Name & Address**  
GOVERNMENT PRIMARY HEALTH CENTRE,  
MEKKRIIMANGALAM-609801, MAYILADUTHURAI DISTRICT.

Details of Device Under Calibration (DUC)			
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD3790211RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY


Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE


Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 97	Uncertainty in % ( ± ) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 125	Uncertainty in % ( ± ) 8.0
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 235	Uncertainty in % ( ± ) 7.2

- REMARKS**
1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
  2. The user should determine the suitability of the instrument for its intended use.
  3. The recalibration interval should be determined on the user requirement.
  4. The results stated in this certificate relate only to the item calibrated.
  5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
  6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By  
  
(Calibration Engineer)  
M.DINESH



Authorised Signatory  
  
 Technical Manager  
C.SHANMUGARAJ

Chief Executive



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO:</b>	SBS/CL/23/13925	<b>MECHANICAL</b>	Page No:1 of 1
Issue Date	10-10-2023		
SRF No & Date	SRF/23/00840-0002 & 09-10-2023		
Receipt Date	09-10-2023		
Calibration Date	09-10-2023		
Calibration Due	08-10-2024		

**Customer Name & Address**  
GOVERNMENT PRIMARY HEALTH CENTRE,  
MEKKRIIMANGALAM-609801, MAYILADUTHURAI DISTRICT.

### Details of Device Under Calibration (DUC)

Description : CENTRIFUGE	Make & Model : REMI & C-854/8
Range : 5250 RPM	Sr. No : NA
Resolution : 10 RPM	Identification No : NA
DUC Condition : Satisfactory	Location : LABORATORY

### Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & Q638380	JRPM-CCTR-A&S-2022-0036	03-10-2023

### **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.7	0.3	4.2
2	2000	1998	2	4.2
3	3000	2997	3	4.2

### **REMARKS**


1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
M.DINESH



 Quality Manager  
(C.SIVABALAN)  Chief Executive

## **SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/13926</b>	<b>MEDICAL DEVICES</b>	<b>Page No:1 of 1</b>
Issue Date	10-10-2023	
SRF No & Date	SRF/23/00840-0003 & 09-10-2023	
Receipt Date	09-10-2023	
Calibration Date	09-10-2023	
Calibration Due	08-10-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
MEKKRIIMANGALAM-609801, MAYILADUTHURAI DISTRICT.

**Details of Device Under Calibration (DUC)**

Description : MICROSCOPE	Make & Model : LABOMED & NA
Range : MULTI	Sr. No : 220438768
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY****RESULTS**


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resistance</b> >20MΩ	<b>Measured values in MΩ</b> 97	<b>Uncertainty in % ( ± )</b> 13.92
2	<b>Earth Leakage</b> <5000µAfor B,BF,CF	<b>Measured values in µA</b> 168	<b>Uncertainty in % ( ± )</b> 7.6
3	<b>Enclosure Leakage</b> <500µAfor B,BF,CF	<b>Measured values in µA</b> 241	<b>Uncertainty in % ( ± )</b> 7.2

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
M.DINESH



Authorised Signatory

  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

### SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13927

Page. No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, MEKKIRIMANGALAM-609801, MAYILADUTHURAI DISTRICT.	SRF No.	SRF/23/00840-0004
	SRF Date	09-10-2023
	Date of Receipt	07-10-2023
	Date of Calibration	09-10-2023
	Due Date for Calibration	08-10-2024
	Issue Date	10-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPEPTE F3
Resolution	1µl	Material	PVC
Serial Number	RW12507	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	100	99.95	99.89	99.87	99.86	99.85	99.87	-0.13	0.03	0.47	
		99.86	99.87	99.86	99.87	99.85					
2	500	499.89	499.87	499.89	499.86	499.85	499.85	-0.15	0.04	0.47	
		499.85	499.76	499.85	499.86	499.85					
3	1000	999.78	999.79	999.78	999.77	999.79	999.78	-0.22	0.01	0.47	
		999.75	999.76	999.78	999.78	999.78					

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 M.ADHIBAN



Authorised by:

  
 Quality Manager/Chief Executive  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13928

Page No : 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, MEKKIRIMANGALAM-609801, MAYILADUTHURAI DISTRICT.	SRF No.	SRF/23/00840-0005
	SRF Date	09-10-2023
	Date of Receipt	07-10-2023
	Date of Calibration	09-10-2023
	Due Date for Calibration	08-10-2024
	Issue Date	10-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	10-100µl	Model	FINNPIPETTE F3
Resolution	0.2µl	Material	PVC
Serial Number	RW09877	Operating Range	10-100µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.6	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Standard Used

SI. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSPL22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl											
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)	
1	10.0	9.96	9.95	9.98	9.96	9.97	9.97	-0.03	0.01	0.47	
		9.98	9.96	9.97	9.98	9.95					
2	50.0	49.86	49.85	49.87	49.86	49.85	49.91	-0.09	0.05	0.47	
		49.95	49.96	49.97	49.95	49.95					
3	100.0	99.98	99.97	99.98	99.96	99.98	99.96	-0.04	0.01	0.47	
		99.95	99.96	99.94	99.95	99.96					

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
6. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade 1 as specified in ISO 3696.

Calibrated By,

(Calibration Engineer)  
 M.ADHIBAN



Authorised by:

Quality Manager / Chief Executive  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.