

Lab Name

PT item

Constituent Group



DEPARTMENT OF CLINICAL BIOCHEMISTRY

## CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023

Lab No

Date of Result Entered:

Date of Report Published:

7446

17/05/2023

05/06/2023

SDI

SI.No Analyte Method / Principle Name Analyzer Name No of Participants DV CV SD

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	184.27 mg/dL	0.36	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.73 mg/dL	-0.23	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	108.38 mg/dL	-1.07	0.27
						•				

SDI Range	Interpretation		
Within -1.00 to +1.00	Excellent.		
Within ±1.01 to ±2.00	Good.		
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.		
Beyond ±3.0	Unacceptable performance. Action Signal.		

Govt. Block PHC, Solakkadu

Lyophilized human serum based

Chemistry III

LAB ADDRESS:

GOVL BIOCK PHC, SOIAKKADU
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT.
SOIAKKADU
TAMILNADU637411

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss

Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\*\*\*\* End of Report \*\*\*\*\*\*\*\*

MEDICAL OFFICER
Government Primary Health Center
Solakadu, Kolli Hills,
Namakkal - 637 411.

