



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023

Lab Name: Govt. Block PHC, Solakkadu
Constituent Group: Chemistry III
PT Item: Lyophilized human serum based

Lab No: 7446
Date of Result Entered: 15/04/2023
Date of Report Published: 06/05/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Year Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	125.4 mg/dL	0.92	0.24
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.47 mg/dL	1.77	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	107.2 mg/dL	-0.74	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within +1.01 to +2.00				Good						
Within +2.01 to +3.99				Accept with caution Warning Signal						
Beyond 4.00				Rejection - Retest - Action Point						

LAB ADDRESS:
GOVT. BLOCK PHC, SOLAKKADU
TAMILNADU 637411

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

PCCL
MEDICAL OFFICER
Government Primary Health Center
Solakkadu, Kollu Hills,
Namakkal - 637 411.